

The Mangum™ Micro Delivery Set is a closed system for use with a syringe pump for administration of IV medications and IV solutions.

This quick reference is provided with the compliments of Smiths Medical, and is intended for use solely as an aid for the participant in developing skill in the use of the product described above. Smiths Medical disclaims responsibility for any use, misuse, or application of this form and makes no representations or warranties concerning competency, proficiency, or skill level obtained or obtainable by the user of the Mangum™ sets. For detailed instructions, specifications, warnings and additional information on the Mangum™ sets or the Medfusion® pump, please refer to the IFU's/Operator's Manual.

Smiths Medical ASD, Inc.
St. Paul, MN 55112, USA
Phone: 1-214-618-0218
Toll-Free USA: 1-800-258-5361

www.smiths-medical.com
Find your local contact information at
www.smiths-medical.com/customer-support




EC Authorized Representative
Smiths Medical International Ltd.
TN25 4BF, UK
Phone: +44 (0) 1233 722100



Mangum™ Micro Delivery Set Quick Reference

Mangum™ Set



Using aseptic technique:

1. Remove sterile product from pouch. Verify that all connections are secure. Close pinch clamp.
2. Close vent on the tubing if spiking a fluid bag; open the vent if spiking a bottle.
3. Remove spike protector and insert spike in the bag or bottle. **(Diagram A)**
4. Remove the protector from extension tubing, attach needle-free access site, and connect syringe to the needle-free access site. **(Diagram B)**
5. Turn stopcock OFF to the patient line. **(Position 1)** Open the pinch clamp and fill syringe with solution from bag or bottle.
6. Turn stopcock OFF to the solution bag or bottle. **(Position 2)** Close pinch clamp.
7. Without disconnecting the syringe, expel air out of the syringe while manually priming the line to the patient with fluid. **(Diagram C)** Tubing is NOT connected to the patient during this process.
8. Load syringe on syringe pump and follow the appropriate syringe pump instructions for fluid or medication administration.

9. Upon completion of fluid or medication administration, turn stopcock OFF to the patient. **(Position 1)** For next infusion, clean and disinfect needle-free access site and connect a new syringe. *(Refer to hospital policy)*

