

CADD®-Solis Ambulatory Infusion Pump Skills Training Checklist for Healthcare Personnel

1. DISCUSS INDICATIONS FOR USE

2. PUMP DESCRIPTION AND BASIC OPERATION

- Pump tour
 - Battery compartment
 - Indicator lights
 - AC/USB/Remote dose cord jacks
 - Power button
 - Latch/lock
 - Air detector
- Pump screens
 - Home screen
 - Status bar
 - Soft key bar
 - Menu screen
 - Protocol title bar
 - Blue help text
- Basic tasks
 - Power pump on/off
 - Battery operations
 - Latch open/close
 - Bag/reservoir change
 - Tubing change
 - Reset reservoir volume
 - Stop/start infusion
- Security codes (3 levels)

3. PUMP PROGRAMMING

- Programming
 - Start new patient
 - Start new protocol/new patient
- Review Intermittent Bolus parameters (if applicable)

Note: When a new library has been sent to the pump, the new library will go into effect the next time the pump is programmed with a new protocol.

4. EDIT/VIEW CURRENT PUMP PROGRAMMING

- Edit delivery settings
- Review/Accept edits to delivery settings (press Start to prompt the Review Settings screen)

5. PUMP TASKS

- Give Clinician Bolus
- Prime tubing
- Set Time and Date
- Adjust Backlight Intensity
- Adjust Alarm volume
- View Reports
- Turn Wireless On/Off *(Available with v4 pump)*
- Adjust Admin Settings

6. ALARMS

- Describe differences between alarm priorities
- Differentiate between "acknowledge" and "silence"
- Identify and resolve alarm conditions
- Alarm help screens

7. REPORTS

View and/or clear reports, as applicable:

- Intermittent Bolus Status
- Given and PCA/PCEA Dose Counters
- PCA/PCEA Dose Graph
- Intermittent Bolus Graph
- Delivery History and Pie Chart
- Current Profile/Protocol *(Available with v4 pump)*
- Delivery Log
- Event Log
- Daylight Saving Time *(Available with v4 pump)*
- Protocol Library Summary
- Wireless Status *(Available with v4 pump)*
- Wireless Settings *(Available with v4 pump)*
- Device Information

This checklist is provided compliments of Smiths Medical and is intended for use solely as an aid for the participant in developing skills in the use of the product described above. Smiths Medical disclaims responsibility for any use, misuse, or application of this form and makes no representations or warranties concerning competency, proficiency, or skill level obtained or obtainable by the user.

Participant (please print): _____

Signature: _____ Date: _____

Instructor signature: _____

PRODUCT(S) DESCRIBED MAY NOT BE LICENSED OR AVAILABLE FOR SALE IN CANADA AND OTHER COUNTRIES

Smiths Medical ASD, Inc.
6000 Nathan Lane North
Minneapolis, MN 55442, USA
Tel: 1-614-210-7300
Toll-Free USA: 1-800-258-5361

Find your local contact information at: www.smiths-medical.com/customer-support

Smiths Medical is part of the global technology business Smiths Group plc. Please see the Instructions for Use/Operator's Manual for a complete listing of the indications, contraindications, warnings and precautions. CADD and the Smiths Medical design mark are trademarks of Smiths Medical. The symbol ® indicates the trademark is registered in the U.S. Patent and Trademark Office and certain other countries. All other names and marks mentioned are the trademarks or service marks of their respective owners. Product referenced is CE-marked.
©2017 Smiths Medical. All rights reserved. IN193537EN-072017

MPAUCA-1661

**Rx
ONLY**

smiths medical