



Smiths Medical is now part of ICU Medical.

Smiths Medical Apache Log4j Software Vulnerability Notice

January 6, 2022

Smiths Medical is aware of the recently announced vulnerability in certain versions of Apache Log4j software identified as [CVE-2021-44228](#). Log4j is a Java-based library to support logging of events in an application and is used in a wide range of enterprise applications. The U.S. Cybersecurity and Infrastructure Security Agency (CISA) has [described](#) this as a critical vulnerability that is being actively exploited. Since the initial announcement of CVE-2021-44228, several other related vulnerabilities have been announced in Log4j, including CVE-2021-45046, CVE-2021-4104 and CVE-45105.

Smiths Medical has investigated and determined that the following products do not contain Log4j software and are not affected by these Log4j vulnerabilities:

Software:

- PharmGuard® Server Infusion Management Software
- PharmGuard Device Reports Software
- PharmGuard Interoperability Software
- Medfusion® 4000 Pump Download Utility (DL4000)
- Medfusion 3500 Pump Download Utility (DL3500)
- PharmGuard Toolbox 2 Software
- PharmGuard Administrator Medication Safety Software
- CADD®-Solis Download Software
- CADD-Solis Network Configuration Utility Software

Infusion Pumps:

- Medfusion 4000 Syringe Infusion Pump
- Medfusion 3500 Syringe Infusion Pump
- CADD-Solis Ambulatory Infusion Pump Model 2110
- CADD-Solis Communication Module Model 2130
- CADD-Solis Communication Module Model 2131

Smiths Medical will continue to closely monitor developments related to the Log4j vulnerabilities and will provide updates to this communication as necessary.

If you have questions about this notification, please contact Smiths Medical Application Support at 1 (800) 258 5361 or ApplicationSupport@smiths-medical.com.